## State of California AIR RESOURCES BOARD

## **EXECUTIVE ORDER I-08-026**

Relating to Independent Contractor Approval under Section 91207, Title 17, California Code of Regulations

BR Laboratories, Incorporated

WHEREAS, the Air Resources Board (ARB), pursuant to section 41512 of the California Health and Safety Code, has established the procedures contained in sections 91200-91220, title 17, California Code of Regulations, to allow the use of independent testers for compliance tests required by ARB;

WHEREAS, it has been determined that BR Laboratories, Incorporated meets the requirements of ARB for conducting ARB Test Methods 1, 2, 3, 4, and 100 (CO, CO<sub>2</sub>, NO<sub>x</sub>, O<sub>2</sub>, THC) pursuant to sections 91200-91220, title 17, California Code of Regulations, when the following condition is met:

1. BR Laboratories, Incorporated uses a 40 percent H<sub>2</sub>/60 percent He or a 40 percent H<sub>2</sub>/60 percent N<sub>2</sub> blend, as appropriate, as fuel for its hydrocarbon analyzer;

WHEREAS, ARB's Executive Officer, pursuant to Health and Safety Code section 39516, issued Executive Order G-02-008, delegating to the Chief of ARB's Monitoring and Laboratory Division the authority to approve independent testers in accordance with title 17, California Code of Regulations, sections 91200-91220.

NOW, THEREFORE, I, William V. Loscutoff, Chief of ARB Monitoring and Laboratory Division, order that BR Laboratories, Incorporated is granted an approval, from the date of execution of this order until June 30, 2009, to conduct the tests listed above, subject to compliance with sections 91200-91220, title 17, California Code of Regulations;

BE IT FURTHER ORDERED that, during the approved period, the Executive Officer or his authorized representative may field audit one or more tests conducted pursuant to this order for each type of testing listed above.

Executed at Sacramento, California this 16<sup>th</sup> day of May 2008

Original signed by

William V. Loscutoff, Chief Monitoring and Laboratory Division